

**Test Report** SL52025245694301TX **Date:** April 27, 2020 **Page 1 of 3**  
HOLLEE FASHIONAL CO.,LTD  
14F BUILDING A, ZHONGLIANG PLAZA, NO.345 YOUYI NORTH STREET, SHIJIAZHUANG, CHINA.

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A)PE gown material 43gsm+2gsm

Composition : (A)polyethylene

Sample Color : (A)blue

End Use : (A)PE gown

Proposed Care Instruction :



Sample Receiving Date : Apr 24, 2020

Testing Period : Apr 24, 2020 - Apr 27, 2020

Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the sample(s) tested, for further details, please refer to the following page(s).

Test Performed : Selected test(s) as requested by applicant



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

Test Report

SL52025245694301TX


Date: April 27, 2020

Page 2 of 3

Conclusion	A	Remark
Water Resistance(Hydrostatic Head)	M	
Water Resistance(Impact Penetration Test)	M	

Remark(s) : M=Meet Client's Requirement

Signed for and on behalf of  
SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center



Sara Guo (Account Executive)

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)



SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd.  
Testing Center

3<sup>rd</sup> Building, No.889, Yishan Road, Xuhui District Shanghai, China 200233  
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233

t (86-21) 61402666 f (86-21) 64958763  
t (86-21) 61402666 f (86-21) 64958763

www.sgs.com.cn  
e [sgs.china@sgs.com](mailto:sgs.china@sgs.com)

**Water Resistance(Hydrostatic Head)**

(AATCC TM 127-2018;Hydrostatic Head; Rate of increase of water pressure: 60 cm H<sub>2</sub>O/min, temp. of distilled water: 21 °C, {Test Side})

As Received

(A)	No. 1	No. 2	No. 3	Average	Requirement
Water Column(cmH <sub>2</sub> O)	>50.0	>50.0	>50.0	>50.0	>50cmH <sub>2</sub> O

**Water Resistance(Impact Penetration Test)**

(AATCC TM 42-2017;Impact Penetration Type I Tester —Tested under controlled condition, water temperature : 27±1 °C)

As Received

(A)	No. 1	No. 2	No. 3	Average	Requirement
The weight of the blotter gained(g)	0.0	0.0	0.0	0.0	<1.0g

Sample Photo



The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

\*\*\*End of Report\*\*\*

